



# 校准证书

## Calibration Certificate



证书编号: 3XJ21700532-0001  
Cer.No.

委托单位: 北京安定医院  
Customer

委托单位地址: 北京市西城区安康胡同5号  
Customer Addr.

仪器名称: 医用离心机  
Description

型号规格: BY-300C  
Model/Type

制造厂商: 北京白洋医疗器械有限公司  
Manufacturer

机身编号: 1807120256  
Serial No.

管理编号: /  
Asset No.

接收日期: 2021-11-09  
Acce. Date

校准日期: 2021-11-09  
Cal. Date

校准人: 张若生  
Calibrated by

签发日期: 2021-11-12  
App. Date

核验人: 赵公卿  
Inspected by

签发人: 赵茂林  
Approved by

印章:  
Chop

地址/ADD: 北京市通州区云杉南路17号  
No.17 Yunshan South Road, Tongzhou District,  
Beijing  
邮编/Post: 101100

电话/Tel: 010-8084 5844  
传真/Fax: 010-8084 5843

网址/website: <http://www.bjlangbo.com>  
邮箱/Email: [bjlangbo@126.com](mailto:bjlangbo@126.com)



# 校准说明

## Directions of Calibration



证书编号: 3XJ21700532-0001  
Certificate No.

- 1. 本实验室质量管理体系符合ISO/IEC 17025:2017《检测和校准实验室能力的通用要求》的要求。**  
The Quality management system of Our Lab meets the ISO/IEC 17025:2017 《General Requirements for the Competence of Testing and Calibration Lab》.
- 2. 本证书未经本实验室书面授权, 不得部分复制。**  
The certificate shall not be partly reproduced without written approval of the laboratory.
- 3. 本次校准结果仅与被校件有关。**  
The calibration results are only related to the items calibrated.
- 4. 证书内页中"P"、"合格"代表"测量结果在允许范围内", "F"、"不合格"代表"测量结果不在允许范围内", "N/A"代表"不适用"; 本证书报告的判定规则和结论仅供参考, 使用人员应结合实际测量的要求合理使用。**  
"P" and "Pass" in this certificate stand for "Low Limit  $\leq$  the measured value  $\leq$  High Limit", "F" and "Fail" Stand for "the measured value  $<$  Low Limit or the measured value  $>$  High Limit", "N/A" stand for "Not Applicable"; The judgment rules and conclusions of this certificate are for reference only, User should use them reasnably according to the actual measurement requirements.
- 5. 环境条件 (Environmental condition):**  
温度(Temperature): 22.3 (°C)      相对湿度(Relative Humidity): 46 (%)
- 6. 校准地点 (The place of the calibration):**  
北京安定医院(疾病样本资源库-实验室)
- 7. 本次校准的技术依据 (Reference documents for the calibration):**  
参照JJF(浙)1117-2015 医用离心机校准规范 Refer to Calibration Specification for Medical Centrifuge
- 8. 本次校准的CNAS认可能力范围及说明 (CNAS accredited scopes and comments).** 详细认可范围请查看CNAS网站中注册号为L9885的证书附件。(Please see the attachment of certificate No.L9885 at CNAS website).

**9. 本次校准所使用的主要测量标准 (Main measurement standards used during the calibration):**

名称 (Description)	出厂编号 (Serial No.)	证书编号 (Certificate No.)	有效期至 (Due Date)	溯源单位 (Traceability to)	技术指标 (Specification)
非接触式转速计	C192469999	1ZJ21700023-0018	2022-08-16	北京朗搏	Urel=0.01%(k=2)

**结论(Conclusion):** 所校准项目合格(Passed at Calibration Items)

**建议校准周期 (Proposed Cal.Period):** 1年/Year



**校准结果(Calibration Results):**

1、外观及功能检查(Apperance and Function Check)

结论  
(Pass/Fail)  
P

2、转速示值误差(Revolution Speed Indication Error)

标称值 (Nominal) (rpm)	标准值 (Reference) (rpm)	误差 (Error) (%)	允许误差 (Limit) (%)	结论 (Pass/Fail)	扩展不确定度 (Urel, k=2) (%)
1000	993	0.70	±2.5	P	0.1
2000	1992	0.40	±2.5	P	0.1
4000	3989	0.28	±2.5	P	0.1

3、转速稳定度(Rotate Speed Stability):

设定值 (Setting) (rpm)	稳定度 (Stability) (%)	允许误差 (Limit) (%)	结论 (Pass/Fail) (P/F)
2000	0.6	≤1.0	P

附(Appendix):

关于测量结果不确定度的说明

(Directions of measurement uncertainty in the calibration)

依据: JJF 1059.1-2012 测量不确定度评定与表示

(Reference document: JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Mearsurement)

-----以下空白(hereafter no data)-----

