

北京市计量检测科学研究院

Beijing Institute of Metrology



21013879003-11-D1971

校准证书

Calibration Certificate

证书编号:

Certificate No. JA21Z-AA100248

客户名称

Client

首都医科大学附属北京安定医院

物品名称

Name of Item

体检秤

型号 / 规格

Model / Specification

HNH-219

物品编号

Serial No.

SK-18A0928

制造单位

Manufacturer

欧姆龙健康医疗(中国)有限公司

客户地址

Address of the Client

北京市西城区德胜门外安康胡同5号

物品接收/校准日期

Date of Receipt/Calibration

2021-05-21/2021-05-21

批准人

Approved by

核验员

Checked by

校准员

Calibrated by



签发日期

Date of Issue

2021

年

06

月

04

日

计量检定机构授权证书号:

Authorization Certificate No.

(国)法计(2017)01007号

地址:

北京市朝阳区安苑东里一区12号

Address:

No.12, 1 block, Anyuandongli, Chaoyang District, Beijing

电话(Tel):

400-090-0026

传真(Fax):

010-57521500

邮编(Post Code):

100029

电子信箱(E-mail):

jly@bjil.cn

北京市计量检测科学研究院校准证书
Calibration Certificate of BIM

证书编号: JA21Z-AA100248 第 2 页共 3 页
Certificate No. Page 2 of 3

北京市计量检测科学研究院是国家法定计量检定机构，检定和校准资格获得国家市场监督管理总局授权。

Beijing Institute of Metrology(BIM) is a National Legal Metrological Verification Institution, which is authorized to conduct verification and calibration services by SAMR.

校准依据 Calibration Reference	JJG 539-2016数字指示秤检定规程				
校准地点 Calibration Location	西城区德胜门外安康胡同5号				
环境条件 Environmental conditions	温度(℃) temperature(℃)	22.6	相对湿度(%) Relative humidity(%)	42	其它 Other

本次校准所使用的主要计量标准器
Measurement Standards for Calibration

名称 Name	测量范围 Measuring Range	不确定度/准确度等级 /最大允许误差 Uncertainty /Accuracy Class/MPE	证书编号/溯源单位 Certificate No./Trace to	有效期至 Date of Expiry
砝码	5kg~2430kg	M ₁ 等级	JA20J- AA001368/北京 计量院	2021/08/05
砝码	100mg~28891g	M ₁ 等级	JA20J- AA001369/北京 计量院	2021/08/05
砝码	500kg~48000kg	M ₁₂ 等级	JA20J- AA001880/北京 计量院	2021/09/24
砝码	1000kg~14000kg	M ₁₂ 等级	JA20J- AA001879/北京 计量院	2021/09/24

- 本证书以中英文两种语言表达，准确含义以中文为准。
The certificate is reported in both English and Chinese, with the Chinese version as standard.
- 本证书未加盖校准专用章无效
The certificate is invalid without the calibration seal of BIM.
- 被校物品修理后，应立即重新校准。
Recalibrate the item as soon as it is repaired.
- 在使用过程中，如对被校物品的技术指标产生怀疑，请重新校准。
Recalibrate the item if there are any doubts about its performance.
- 根据顾客的要求，复校时间间隔为(/)个月。
According to requirements of the client, the calibration interval should be (/) months.

未经本院书面批准，不得部分复制本证书。
This certificate shall not be reproduced except in full, without written approval of BIM.

北京市计量检测科学研究所校准证书

Calibration Certificate of BIM

证书编号:

JA21Z-AA100248

Certificate No.

第

3

页共

3

页

Page 3 of 3

校准结果

Results of Calibration

1. 称量测试 (kg)

化整前示值 (kg)

备注

20

19.9

Max=200kg

40

39.8

d=100g

60

59.8

80

80.1

校准点:80kg

100

100.2

120

120.2

200

200.2

2. 重复性测试 (kg)

$P_{Max}-P_{min}$ (g)

80

100

本次校准结果的扩展不确定度为

$U = 104g$ ($k=2$)

以下空白(End)

证书骑缝专用章(1)

未经本院书面批准, 不得部分复制本证书。

This certificate shall not be reproduced except in full, without written approval of BIM.