

# 北京市计量检测科学研究院

Beijing Institute of Metrology



21013855002-11-B117

## 校准证书

Calibration Certificate

证书编号:  
Certificate No. JC21Z-AB100328

客户名称

Client

首都医科大学附属北京安定医院

物品名称

Name of Item

体检秤

型号 / 规格

Model / Specification

HNH-219

物品编号

Serial No.

SK-18A0928

制造单位

Manufacturer

欧姆龙健康医疗(中国)有限公司

客户地址

Address of the Client

北京市西城区德胜门外安康胡同5号

物品接收/校准日期

Date of Receipt/Calibration

2021-05-21/2021-05-21

批准人

Approved by

核验员

Checked by

校准员

Calibrated by



签发日期

Date of Issue

2021

年

06

月

07

日

计量检定机构授权证书号:

Authorization Certificate No.

(国)法计(2017)01007号

地址:

Address:

北京市朝阳区安苑东里一区12号

No.12, 1 block, Anyuandongli, Chaoyang District, Beijing

电话(Tel):

传真(Fax):

邮编(Post Code):

电子信箱(E-mail):

400-090-0026

010-57521500

100029

jly@bjil.cn

## 北京市计量检测科学研究所校准证书

Calibration Certificate of BIM

第 2 页共 3 页

证书编号: JC21Z-AB100328  
Certificate No.

Page 2 of 3

北京市计量检测科学研究所是国家法定计量检定机构, 检定和校准资格获得国家市场监督管理总局授权。

Beijing Institute of Metrology(BIM) is a National Legal Metrological Verification Institution, which is authorized to conduct verification and calibration services by SAMR.

|                                  |                                             |      |                                 |    |               |
|----------------------------------|---------------------------------------------|------|---------------------------------|----|---------------|
| 校准依据<br>Calibration Reference    | JJG 1-1999钢直尺检定规程<br>JJG 928-1998超声波测距仪检定规程 |      |                                 |    |               |
| 校准地点<br>Calibration Location     | 首都医科大学附属北京安定医院                              |      |                                 |    |               |
| 环境条件<br>Environmental conditions | 温度(°C)<br>temperature(°C)                   | 20.0 | 相对湿度(%)<br>Relative humidity(%) | 45 | 其它<br>Other / |

本次校准所使用的主要计量标准器  
Measurement Standards for Calibration

| 名称<br>Name | 测量范围<br>Measuring Range | 不确定度/准确度等级<br>/最大允许误差<br>Uncertainty /Accuracy Class/MPE | 证书编号/溯源单位<br>Certificate No./Trace to | 有效期至<br>Date of Expiry |
|------------|-------------------------|----------------------------------------------------------|---------------------------------------|------------------------|
| 手持式激光测距仪   | (0~45)m                 | $U = (0.05 + 0.03L) \text{ mm } (k=2)$<br>(L-m)          | CDJx2020-00862/中国计量院                  | 2022/04/14             |
|            |                         |                                                          |                                       |                        |
|            |                         |                                                          |                                       |                        |
|            |                         |                                                          |                                       |                        |
|            |                         |                                                          |                                       |                        |

- ☐ 本证书以中英文两种语言表达, 准确含义以中文为准。  
The certificate is reported in both English and Chinese, with the Chinese version as standard.
- ☐ 本证书未加盖校准专用章无效  
The certificate is invalid without the calibration seal of BIM.
- ☐ 被校物品修理后, 应立即重新校准。  
Recalibrate the item as soon as it is repaired.
- ☐ 在使用过程中, 如对被校物品的技术指标产生怀疑, 请重新校准。  
Recalibrate the item if there are any doubts about its performance.
- ☐ 根据顾客的要求, 复校时间间隔为( / )个月。  
According to requirements of the client, the calibration interval should be ( / ) months.

未经本院书面批准, 不得部分复制本证书。

This certificate shall not be reproduced except in full, without written approval of BIM.

北京市计量检测科学研院校准证书  
Calibration Certificate of BIM

证书编号: JC21Z-AB100328  
Certificate No.

第 3 页共 3 页  
Page 3 of 3

校 准 结 果  
Results of Calibration

| 标准高度 (cm) | 仪器 示值 (cm) |
|-----------|------------|
| 80.0      | 80.0       |
| 120.0     | 120.0      |
| 160.0     | 160.0      |
| 200.0     | 200.0      |

在(80.0~200.0)cm范围内, 示值误差不超过 $\pm 0.5$ cm范围。

测量结果不确定度:  $U=0.3\text{cm}$   $k=2$

以下空白 (End)



证书骑缝专用章(1)

未经本院书面批准, 不得部分复制本证书。

This certificate shall not be reproduced except in full, without written approval of BIM.